




## CERTIFICATE OF ANALYSIS

PO BOX 130549 Spring, TX 77393  
Phone: (281) 298-9410 Fax: (281) 298-9411

<p><b>FINISHED PRODUCT, LOT NUMBER, MFG. /EXP DATE:</b> PolySeed® • Part No. P-110 • Lot 042006 • Mfg. Date: 06/2020 • Exp. Date: 06/2022</p>
<p><b>FORMULATION:</b> The formulation for this product contains a range of naturally occurring microorganisms, which are known to be non-pathogenic to man or animals.</p>
<p><b>VIABLE COUNT, FINAL TEST RESULT:</b> The product has been fully tested in accordance with Finished Product Specifications and contains a minimum viable count of <math>4.00 \times 10^9</math> cfu/g</p>
<p><b>GLUCOSE/GLUTAMIC-ACID RESULTS:</b> Tested results within acceptable range 198 +/- 30.5 mg/L (167.5 - 228.5 mg/L). GGA Lot# J172-23 – Average Test Result: 215  See <a href="http://www.polyseed.com">www.polyseed.com</a> for details</p>
<p><b>SEED CONTROL FACTOR:</b> Tested results within acceptable range 0.6 – 1.0 see <a href="http://www.polyseed.com">www.polyseed.com</a> for details</p>
<p><b>SALMONELLA TEST RESULT:</b> The product has been shown to be Salmonella negative using procedures recommended in the Microbiology Laboratory Guidebook, published by the USDA Food Safety and Inspection Service.</p>
<p>The purpose of this document is to assure that the Finished Product conforms to the above specification.</p> <p style="text-align: center;">   <b>Signature:</b> _____ <b>Date:</b> 06/16/2020  <i>Quality Control Department</i> </p>
<p>POLYSEED.Ref.1.19 <span style="float: right;">Revised Jan 20</span></p>